


INFORMATIVE NOTE
Name: REFLEX 51
Date: 11/05/2021 Rev: 03
MANUFACTURER

 REFLEX SpA
 Via Passeri, 2 46019 VIADANA (MN)
 Tel.: + 39 0375 758891 Fax + 39 0375 464504
info@reflexx.com
www.reflexx.com

Pack

| Cod./Size | 100 | |
|-----------|---------------|---------------|
| | Box | 10 Carton |
| R51/XS | 8032891632994 | 8032891637999 |
| R51/S | 8032891633007 | 8032891638002 |
| R51/M | 8032891633014 | 8032891638019 |
| R51/L | 8032891633021 | 8032891638026 |
| R51/XL | 8032891633038 | 8032891638033 |

DESCRIPTION: ■ disposable examination glove powdered, not sterile ■ **Material:** natural latex ■ **Shape:** ambidextrous; five fingers, beaded cuff ■ **Colour:** white ■ **Feature:** smooth surface ■ **Packing:** 100 pieces/box; 10 boxes/carton ■ **ISO 2859:** AQL 1.5 (Level G1) freedom from holes; AQL 2.5 (level G1) major defects; AQL 4 (level G1) minor

CONFORMITY AND PERFORMANCES
Personal Protective Equipment cat III

0465 ANCI_CIMAC, Via Aguzzafame, 60/b 27029 Vigevano (PV)

Reg EU 2016/425 - EN ISO 374-1:2016 TYPE C - EN ISO 374-5:2016 - EN 374-2:2014 - EN 420:2003 + A1 2009

Dispositivo Medico

Reg EU 2017/745 -EN 455 1,2,3 & 4

(CND T010201 NON SURGICAL LATEX GLOVES)



Reg CE 1935/2004 – Reg EU 10/2011 - DM 21/03/19. To such use carefully read the declaration of suitability for food contact (DICA).

DIMENSIONS

| Cod./Size | Length. min(mm) | Width. (±5 mm) |
|--------------|-----------------|----------------|
| R51/ XS (6) | 240 | 75 |
| R51/ S (7) | 240 | 85 |
| R51/ M (8) | 240 | 95 |
| R51/ L (9) | 240 | 106 |
| R51/ XL (10) | 240 | 116 |


| Thickness mm | Palm (+/- 0.02) | 0.08 |
|--------------|-----------------|------|
|--------------|-----------------|------|

| EN 455-2:2009/A1 | Tensile strength | ≥6 N |
|------------------|------------------|------|
|------------------|------------------|------|

| EN ISO 374-5:2016 | Protection from microorganisms (bacteria and fungi) | Virus protection ISO 16604:2004 |
|-------------------|---|---------------------------------|
| | pass | pass |

EN ISO 374-1:2016

| Chemical compound: | EN 16523-1:2015 TYPE C | | | EN 374-4:2013 |
|----------------------|------------------------|-------------------|-------------------|---------------|
| | COD | Performance level | Breakthrough time | % degradation |
| Sodium hydroxide 40% | K | 2 | >30' | -0.8% |

REG EU 2016/425

EN 420:2003 + A1:2009

RISKS FROM WHICH THE DPI IS DESIGNED TO PROTECT:

| COD | Protection |
|------------------------|---|
| K Sodium hydroxide 40% | CAS:1310-73-2 Protection from use of inorganic bases |

RECOMENDATIONS OF USE Multi-purpose glove to protect the hand and/or the handled product, can also be used for general uses ■ The glove is suitable for contact with pure chemical products for a period of less than the permeation time indicated in the chemical resistance table ■ Results obtained under laboratory conditions may not reflect actual working time **WARNINGS** Before using the product, try it in the real condition of use ■ The glove is a device subject to wear, so frequent

REFLEX S.p.A. Unipersonale Via Passeri, 2 - 46019 Viadana (MN) | Tel. + 39 0375 758891 Fax + 39 0375 464504

 Capitale Sociale 480.000€ (i.v.) | P. IVA 02085450209 | R.E.A. 223166 | info@reflexx.com www.reflexx.com


replacement is recommended and in all cases where imperfections are observed ■ Wear gloves with dry, clean hands ■ The product contains natural latex which can cause allergic reactions of type I and IV hypersensitivity. In case of reaction, discontinue use and consult a doctor ■ **STORAGE AND DISPOSAL** Store the gloves in their original packaging in a dry place. ■ Do not expose to direct sunlight and heat sources. ■ The expiry date is 3 years from the date of manufacture indicated on the sales package, and is valid for the product properly stored ■ Gloves must be disposed of according to local regulations ■ The user must inform the manufacturer and the relative authority of any incidents occurring during the use of the protective equipment ■ Declaration of Conformity on www.reflexx.com